

Delft University of Technology

### **Thrombus mechanics**

## How can we contribute to improve diagnostics and treatment?

Migliavacca, Francesco; Luraghi, Giulia; Akyildiz, Ali C.; Gijsen, Frank J.H.

DOI 10.1016/j.jbiomech.2021.110935

Publication date 2022 **Document Version** Final published version

Published in Journal of Biomechanics

**Citation (APA)** Migliavacca, F., Luraghi, G., Akyildiz, A. C., & Gijsen, F. J. H. (2022). Thrombus mechanics: How can we contribute to improve diagnostics and treatment? *Journal of Biomechanics*, *132*, Article 110935. https://doi.org/10.1016/j.jbiomech.2021.110935

#### Important note

To cite this publication, please use the final published version (if applicable). Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights. We will remove access to the work immediately and investigate your claim.

# Green Open Access added to TU Delft Institutional Repository

# 'You share, we take care!' - Taverne project

https://www.openaccess.nl/en/you-share-we-take-care

Otherwise as indicated in the copyright section: the publisher is the copyright holder of this work and the author uses the Dutch legislation to make this work public. ELSEVIER

Contents lists available at ScienceDirect

# Journal of Biomechanics

journal homepage: www.elsevier.com/locate/jbiomech



### Editorial

### Thrombus mechanics: How can we contribute to improve diagnostics and treatment?



This special issue aims to introduce the reader the most recent advances in the area of stroke treatment with an eye on the mechanical and modeling aspects.

Treatment of acute ischemic stroke (AIS) has gained interest in the recent years with the introduction, first of intravenous administration of thrombus-dissolving drugs, and more recently of endovascular thrombectomy (EVT) with stent retrievers and aspiration devices or a combination of the two. Mechanical aspects play a crucial role in various stages of the disease. The review by *Ospel* and colleagues (*Ospel* et al., 2021) introduces the reader to IAS and the treatment options, from the intravenous thrombolysis to the new EVT as the mechanical thrombectomy with both stent-retrievers and aspiration catheters. The authors provide from a clinical viewpoint a thorough overview of the mechanical factors that influence the EVT success, including factors on blood vessels, thrombus, procedure techniques, the interaction between devices and thrombus and vessels.

AIS research is devoted to study new technologies and treatment options by using in vivo, ex-vivo, in vitro and in silico tests. The important role of ex-vivo and in vitro modeling is comprehensively discussed in the review by McCarthy and colleagues (McCarthy et al., 2021) especially in light of the lack of in vivo measurements related to thrombus composition and shape. From a medical device developers' perspective, it is important to perform controlled experiments to qualitatively and quantitatively investigate clot mechanics. Luraghi and colleagues (Luraghi et al., 2021c) present an overview of the literature within the field of in vitro and in silico evaluation of EVT procedure. They categorized the studies based on the vessel, clot, device and procedure characteristics, geometries, and models. The use of animal and in vitro experiments is discussed in the work by Gounis and colleagues (Anagnostakou et al., 2022), who analyzed pre-clinical surrogate models of AIS treatment. They highlight the importance of the transition from preclinical tests to clinical applications as the former presents limitations, especially in satisfying the criteria established by the regulatory agencies. In this view, validation is usually proposed for a specific context of use and used to build credibility to the numerical model based on the ASME V&V40 protocol. The credibility refers to the level of trust of a computational test with respect to the context of the model use and is paramount when applying the results in safety-critical situations, as for the EVT application. Luraghi and colleagues (Luraghi et al., 2021a) proposed an applicability evaluation of a numerical thrombectomy model using a rigorous step-by-step method. Focusing on the credibility of the preclinical tests, they demonstrated the reliability and validity of an in silico thrombectomy model based on validation evidence.

First and foremost, the mechanical properties of the thrombus under various loading conditions and how these are related to thrombus

Available online 23 December 2021 0021-9290/© 2021 Published by Elsevier Ltd. composition need to be established since they are essential data to better investigate optimization of AIS treatment. A novel computational model of microvascular thrombus formation is proposed by Masalceva and colleagues (Masalceva et al., 2022), based on 2D particle-based model of thrombus shell formation combined with platelet agonists kinetics. The strength of their model lies in its ability in analyzing the distribution of inter-platelet forces throughout the thrombus by coupling local biochemical cues to the inter-platelet interactions and granule secretion. Their study revealed multiple weak spots in the outer regions of thrombus, important issue in the thrombus formation process. On the same path, Rezaeimoghaddam and colleagues (Rezaeimoghaddam and Van de Vosse, 2022) present a mathematical model amenable to predict thrombosis and identify the risk of thrombus formation when devices are used. The developed method, which combines computational fluid dynamics with a set of convection-diffusion-reaction equations, is also compared to in vitro experimental results, aimining at identifying the most relevant equations by understanding signaling pathways. Teeraratkul and colleagues (Teeraratkul and Mukherjee, 2021) focus on the important phenomena of flow-mediated transport of biochemical species. They developed a microstructure model with a custom preconditioned fictitious domain formulation for thrombus-hemodynamics interactions and an advection-diffusion formulation for transport. They amied at demonstrating the formation of coherent flow structures around the thrombus which organize advective transport which plays an important role in the fundamental processes underlying flow-mediated transport around the thrombus, permeation into the thrombus, and subsequent intra-thrombus transport.

After the thrombus gets formed, the interaction between the thrombus and the blood flow is important to understand the relationship between the origin of the thrombus (left atrium, carotid bifurcation) and the location where the thrombus finally lodges in the intracranial circulation, and the consequences associated with the lodging location. The in vivo study presented by Georgakopoulous and colleagues (Georgakopoulou et al., 2021) analyses the effects of microemboli on the brain tissue. The objective of their study was to determine the extent of damage on brain tissue caused by microemboliby using rat models. They establish that microemboli, which individually may not always be harmful, could interact and result in local hypoxia or infarction in case of aggregation. Xue and colleagues (Xue et al., 2021) propose a numerical model to quantify the effects of cerebral microthrombi on oxygen transport to tissue in terms of hypoxia and ischemia. The implemented Green's function method was used to statistically represent the capillary networks in view of oxygen transport investigation. Quantifying the impact of microthrombi on oxygen transport and tissue death is believed to be extremely important in brain models of ischaemic

stroke and thrombectomy.

When diagnosing patients with intracranial thrombus, the extraction of additional information like the thrombus mechanical properties and perviousness is of fundamental importance. Cahalane and colleagues (Cahalane et al., 2021), in their review study, stressed particularly the usefulness and feasibility to quantitatively assess compositional and mechanical properties of thombi with imaging. They present a scoping review detailing associations between thrombus composition, mechanical behaviour and radiological imaging characteristics. They point out the importance of correctly identifying with imaging techniques if a thrombus is fibrin-rich or red blood cell-rich, which have different mechanical properties, to choose the best thrombectomy approach. The work by Santos and colleagues (Santos et al., 2021) studies the link between perviousness detected by thrombus attenuation measurements and functional outcome. They state that thrombus perviousness measurements from entire thrombus segmentation have a stronger impact on functional patient outcome compared to the current clinical approach of manual assessments.

In the majority of the case, the first treatment of AIS involves the administration of intravenous recombinant-tissue plasminogen activator (r-tPA) as soon as the patient arrives at the hospital. A perspective paper on thrombolysis with particular emphasis on numerical modeling is presented by Petkantchin and colleagues (Petkantchin et al., 2022). The authors point out the difficulties to have a complete picture of the thrombolysis problem and they present the steps to build a numerical model describing the evolution of a patient's clot under the action of t-tPA. *Dwivedi* and colleagues (Dwivedi et al., 2021) quantified the effect of r-tPA on clot analogs' mechanical behavior in view of its impact on the mechanical thrombectomy. They concluded that the red blood cell content, platelet retraction and exposure time to r-tPA on the contrary, was not a significant factor in reducing clot mass or changing mechanical properties of the clots.

Today, the preferred treatment of ischemic stroke is the rapid restoration of blood flow by means of EVT. The treatment consists of the mechanical removal of a blood clot by a stent retriever and/or an aspiration catheter. A major shortcoming of EVT is that in 20% of the cases the occluding thrombus cannot be removed from the artery by a stent retriever. Recent studies (Dutra et al., 2019; Ospel et al., 2019; Pérez-García et al., 2020; Zhang et al., 2021) have demonstrated that large fragments of thrombus may be left behind after EVT in stroke patients. This aspect of EVT, which impacts the procedure outcome by incomplete restoration of blood flow, re-occlusion or distal embolization, is presently difficult to predict and control. Incomplete removal of the thrombus is assumed to result from a poor match between the highly variable mechanical properties of the thrombus and the design of the stent retriever. In the paper by Moerman and colleagues (Moerman et al., 2021), a costumized and automated platform for the generation of highresolution patient-specific models of cerebral arteries directly from clinical images was developed. It is the first step in the pre-processing phase to have a patient-specific virtual representation of the EVT procedure. In the study by Luraghi and colleagues (Luraghi et al., 2021b), the first finite element simulations on patient-specific thrombectomy were carried out by comparing virtual results with real patient outcomes. The simulations showed, as in reality did happen, that two attempts of stent-retriever passes were necessary to remove the clot with no fragmentations to follow.

Finally, the removal of the thrombus is often done by direct aspiration as a first-line option. How the thrombi behave during aspiration, and how they might fail during such a procedure is of crucial importance. *Fereidoonnezhad* and colleagues (Fereidoonnezhad and McGarry, 2021) and *Oyekola* and colleges (Oyekole et al., 2021) investigated the aspiration procedure by means of in silico models. The former mainly focuses on the development of a new constitutive model that incorporates the unrecoverable plastic deformation of clots, calibrated with experiments. The new model which links the plastic deformation and cross-link formation/dissociation in blood clots was used to simulate the aspiration procedure. The latter proposes a viscoelastic solid model and a viscoelastic cohesive interface between the clot and the artery to virtually reproduce the EVT. Their results indicate the likelihood of the clot to be pushed outward against the artery, close to the catheter tip.

From the collection of the papers presented in this special issue, it is clear that mechanics plays a crucial role in the treatment of AIS. It is our hope that deepening our knowledge of thrombus mechanics will help us to optimize treatment strategies and devices.

#### **Declaration of Competing Interest**

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

#### References

- Anagnostakou, V., Epshtein, M., Kühn, A.L., King, R.M., Puri, A., Gounis, M.J., 2022. Preclinical modeling of mechanical thrombectomy. J. Biomech. 130, 110894. https://doi.org/10.1016/j.jbiomech.2021.110894.
- Cahalane, R., Boodt, N., Akyildiz, A.C., Giezen, J.-A., Mondeel, M., van der Lugt, A., Marquering, H., Gijsen, F., 2021. A review on the association of thrombus composition with mechanical and radiological imaging characteristics in acute ischemic stroke. J. Biomech. 129, 110816. https://doi.org/10.1016/j.jbiomech.2021.110816.
- Dutra, B.G., Tolhuisen, M.L., Alves, H.C.B.R., Treurniet, K.M., Kappelhof, M., Yoo, A.J., Jansen, I.G.H., Dippel, D.W.J., van Zwam, W.H., van Oostenbrugge, R.J., da Rocha, A.J., Lingsma, H.F., van der Lugt, A., Roos, Y.B.W.E.M., Marquering, H.A., Majoie, C.B.L.M., MR CLEAN Registry Investigators, 2019. Thrombus imaging characteristics and outcomes in acute ischemic stroke patients undergoing endovascular treatment. Stroke 50, 2057–2064. https://doi.org/10.1161/ STROKEAHA.118.024247.
- Dwivedi, A., Glynn, A., Johnson, S., Duffy, S., Fereidoonnezhad, B., McGarry, P., Gilvarry, M., McCarthy, R., 2021. Measuring the effect of thrombosis, thrombus maturation and thrombolysis on clot mechanical properties in an in-vitro model. J. Biomech. 129, 110731. https://doi.org/10.1016/j.jbiomech.2021.110731.
- Fereidoonnezhad, B., McGarry, P., 2021. A new constitutive model for permanent deformation of blood clots with application to simulation of aspiration thrombectomy. J. Biomech. 130, 110865. https://doi.org/10.1016/j.jbiomech.2021.110865.
- Georgakopoulou, T., van der Wijk, A.-E., Bakker, E.N.T.P., vanBavel, E.d., Majoie, C., Marquering, H., van Bavel, E.d., Hoekstra, A., Dippel, D., Lingsma, H., van der Lugt, A., Samuels, N., Boodt, N., Roos, Y., De Meyer, S.F., Staessens, S., Vandelanotte, S., Konduri, P., Terreros, N.A., Chopard, B., Raynaud, F., Petkantchin, R., Panteleev, M., Shibeko, A., Boudjeltia, K.Z., Blanc-Guillemaud, V., Migliavacca, F., Dubini, G., Luraghi, G., Matas, J.F.R., Bridio, S., Mc Garry, P., Gilvarry, M., McCarthy, R., Moerman, K., Fereidoonnezhad, B., Dwivedi, A., Duffy, S., Payne, S., Jozsa, T., Georgakopoulou, S., Padmos, R., Azizi, V., Miller, C., van der Kolk, M., 2021. Quantitative 3D analysis of tissue damage in a rat model of microembolization. J. Biomech. 128, 110723. https://doi.org/10.1016/j. ibiomech.2021.110723.
- Luraghi, G., Bridio, S., Miller, C., Hoekstra, A., Rodriguez Matas, J.F., Migliavacca, F., 2021a. Applicability analysis to evaluate credibility of an In Silico Thrombectomy procedure. J. Biomech. 126, 110631. https://doi.org/10.1016/j. ibiomech.2021.110631.
- Luraghi, G., Bridio, S., Rodriguez Matas, J.F., Dubini, G., Boodt, N., Gijsen, F.J.H., van der Lugt, A., Fereidoonnezhad, B., Moerman, K.M., McGarry, P., Konduri, P.R., Arrarte Terreros, N., Marquering, H.A., Majoie, C.B.L.M., Migliavacca, F., 2021b. The first virtual patient-specific thrombectomy procedure. J. Biomech. 126, 110622. https://doi.org/10.1016/j.jbiomech.2021.110622.
- Luraghi, G., Cahalane, R.M.E., van de Ven, E., Overschie, S.C.M., Gijsen, F.J.H., Akyildiz, A.C., 2021c. In vitro and in silico modeling of endovascular stroke treatments for acute ischemic stroke. J. Biomech. 127, 110693. https://doi.org/10.1016/ j.jbiomech.2021.110693.
- Masalceva, A.A., Kaneva, V.N., Panteleev, M.A., Ataullakhanov, F., Volpert, V., Afanasyev, I., Nechipurenko, D.Y., 2022. Analysis of microvascular thrombus mechanobiology with a novel particle-based model. J. Biomech. 130, 110801. https://doi.org/10.1016/j.jbiomech.2021.110801.
- McCarthy, R., Mirza, M., Johnson, S., Dwivedi, A., Gunning, G., Vale, D., Gilvarry, M., 2021. Aspects of ischemic stroke biomechanics derived using ex-vivo and in-vitro methods relating to mechanical thrombectomy. J. Biomech. 110900. https://doi. org/10.1016/j.jbiomech.2021.110900.
- Moerman, K.M., Konduri, P.R., Fereidoonnezhad, B., Marquering, H.A., van der Lugt, A., Luraghi, G., Bridio, S., Migliavacca, F., Rodriguez Matas, J.F., McGarry, P., 2021. Development of a patient-specific cerebral vasculature fluid-structure-interaction model. J. Biomech. https://doi.org/10.31224/osf.io/qaujs.
- Ospel, J.M., van der Lugt, A., Gounis, M., Goyal, M., Majoie, C.B.L.M., 2021. A clinical perspective on endovascular stroke treatment biomechanics. J. Biomech. 127, 110694. https://doi.org/10.1016/j.jbiomech.2021.110694.

#### Editorial

- Ospel, J.M., Volny, O., Jayaraman, M., McTaggart, R., Goyal, M., 2019. Optimizing fast first pass complete reperfusion in acute ischemic stroke-the BADDASS approach (BAlloon guibe with large bore Distal Access catheter with dual aspiration with Stent-retriever as Standard approach). Expert Rev. Med. Devices 16 (11), 955–963. https://doi.org/10.1080/17434440.2019.1684263.
- Oyekole, O., Simon, S., Manning, K.B., Costanzo, F., 2021. Modeling acute ischemic stroke recanalization through cyclic aspiration. J. Biomech. 128, 110721. https:// doi.org/10.1016/j.jbiomech.2021.110721.
- Pérez-García, C., Moreu, M., Rosati, S., Simal, P., Egido, J.A., Gomez-Escalonilla, C., Arrazola, J., 2020. Mechanical thrombectomy in medium vessel occlusions: Blind exchange with mini-pinning technique versus mini stent retriever alone. Stroke 51 (11), 3224–3231. https://doi.org/10.1161/STROKEAHA.120.030815.
- Petkantchin, R., Padmos, R., Zouaoui Boudjeltia, K., Raynaud, F., Chopard, B., 2022. Thrombolysis: observations and numerical models. J. Biomech. https://doi.org/ 10.1016/j.jbiomech.2021.110902.

Rezaeimoghaddam, M., Van de Vosse, F., 2022. A continuum model of thrombus formation and development in cardiovascular devices. J. Biomech. https://doi.org/ 10.1016/j.jbiomech.2021.110915.

- Santos, E.M.M., Arrarte Terreros, N., Kappelhof, M., Borst, J., Boers, A.M.M., Lingsma, H. F., Berkhemer, O.A., Dippel, D.W.J., Majoie, C.B., Marquering, H.A., Niessen, W.J., 2021. Associations of thrombus perviousness derived from entire thrombus segmentation with functional outcome in patients with acute ischemic stroke. J. Biomech. 128, 110700. https://doi.org/10.1016/j.jbiomech.2021.110700.
- Teeraratkul, C., Mukherjee, D., 2021. Microstructure aware modeling of biochemical transport in arterial blood clots. J. Biomech. 127, 110692. https://doi.org/10.1016/ j.jbiomech.2021.110692.
- Xue, Y., El-Bouri, W.K., Józsa, T.I., Payne, S.J., 2021. Modelling the effects of cerebral microthrombi on tissue oxygenation and cell death. J. Biomech. 127, 110705. https://doi.org/10.1016/j.jbiomech.2021.110705.
- Zhang, Y., Zhang, Y., Hu, C., Zhao, W., Zhang, Z., Li, W., 2020. A direct aspiration firstpass technique (ADAPT) versus stent retriever for acute ischemic stroke (AIS): a

systematic review and meta-analysis. J. Neurol. 268 (12), 4594–4606. https://doi. org/10.1007/s00415-020-10284-w.

Francesco Migliavacca

Department of Chemistry, Materials and Chemical Engineering "Giulio Natta", Politecnico di Milano, Milan, Italy

Giulia Luraghi Department of Chemistry, Materials and Chemical Engineering "Giulio Natta", Politecnico di Milano, Milan, Italy Department of Biomechanical Engineering, Delft University of Technology, Delft, the Netherlands

Ali C. Akyildiz, Frank J.H. Gijsen Department of Biomechanical Engineering, Delft University of Technology, Delft, the Netherlands Department of Biomedical Engineering, Thoraxcenter, Erasmus Medical Center, Rotterdam, the Netherlands

\* Corresponding author at: Department of Chemistry, Materials and Chemical Engineering 'Giulio Natta' Politecnico di Milano, Piazza L. da Vinci 32, 20133 Milan, Italy. Tel.: +39 02 2399 4316. *E-mail address: francesco.migliavacca@polimi.it* (F. Migliavacca).