

## Raindrop intervalometer

van de Giesen, Nick; Hut, Rolf; ten Veldhuis, Marie-claire

**Publication date**

2017

**Document Version**

Final published version

**Published in**

Geophysical Research Abstracts (online)

**Citation (APA)**

van de Giesen, N., Hut, R., & ten Veldhuis, M. (2017). Raindrop intervalometer. *Geophysical Research Abstracts (online)*, 19, Article EGU2017-3196.

**Important note**

To cite this publication, please use the final published version (if applicable).  
Please check the document version above.

**Copyright**

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

**Takedown policy**

Please contact us and provide details if you believe this document breaches copyrights.  
We will remove access to the work immediately and investigate your claim.



## Raindrop intervalometer

Nicolaas van de Giesen, Rolf Hut, and Marie-claire ten Veldhuis

Delft University of Technology, Water Resources Engineering, Delft, Netherlands (n.c.vandegiesen@tudelft.nl)

If one can assume that drop size distributions can be effectively described by a generalized gamma function [1], one can estimate this function on the basis of the distribution of time intervals between drops hitting a certain area. The arrival of a single drop is relatively easy to measure with simple consumer devices such as cameras or piezoelectric elements. Here we present an open-hardware design for the electronics and statistical processing of an intervalometer that measures time intervals between drop arrivals.

The specific hardware in this case is a piezoelectric element in an appropriate housing, combined with an instrumentation op-amp and an Arduino processor. Although it would not be too difficult to simply register the arrival times of all drops, it is more practical to only report the main statistics. For this purpose, all intervals below a certain threshold during a reporting interval are summed and counted. We also sum the scaled squares, cubes, and fourth powers of the intervals. On the basis of the first four moments, one can estimate the corresponding generalized gamma function and obtain some sense of the accuracy of the underlying assumptions.

Special attention is needed to determine the lower threshold of the drop sizes that can be measured. This minimum size often varies over the area being monitored, such as is the case for piezoelectric elements. We describe a simple method to determine these (distributed) minimal drop sizes and present a bootstrap method to make the necessary corrections.

### Reference

[1] Uijlenhoet, R., and J. N. M. Stricker. "A consistent rainfall parameterization based on the exponential raindrop size distribution." *Journal of Hydrology* 218, no. 3 (1999): 101-127.