

Free-breathing simultaneous T_1 , T_2 , and $T2^*$ quantification in the myocardium

Hermann, Ingo; Kellman, Peter; Demirel, Omer B.; Akçakaya, Mehmet; Schad, Lothar R.; Weingärtner, Sebastian

DOI

[10.1002/mrm.28753](https://doi.org/10.1002/mrm.28753)

Publication date

2021

Document Version

Final published version

Published in

Magnetic Resonance in Medicine

Citation (APA)

Hermann, I., Kellman, P., Demirel, O. B., Akçakaya, M., Schad, L. R., & Weingärtner, S. (2021). Free-breathing simultaneous T_1 , T_2 , and $T2^*$ quantification in the myocardium. *Magnetic Resonance in Medicine*, 86(3), 1226-1240. <https://doi.org/10.1002/mrm.28753>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

Free-breathing simultaneous T_1 , T_2 , and T_2^* quantification in the myocardium

Ingo Hermann^{1,2}  | Peter Kellman³ | Omer B. Demirel⁴ | Mehmet Akçakaya⁴ |
Lothar R. Schad² | Sebastian Weingärtner¹

¹Department of Imaging Physics, Magnetic Resonance Systems Lab, Delft University of Technology, Delft, The Netherlands

²Computer Assisted Clinical Medicine, Medical Faculty Mannheim, Heidelberg University, Mannheim, Germany

³National Heart, Lung, and Blood Institute, National Institutes of Health, DHHS, Bethesda, MD, USA

⁴Department of Electrical and Computer Engineering and Center for Magnetic Resonance Research, University of Minnesota, Minnesota, MN, USA

Correspondence

Ingo Hermann, Delft University of Technology, Department of Imaging Physics, Magnetic Resonance Systems Lab, Lorentzweg 1, 2628 CJ Delft, The Netherlands.

Email: I.Hermann@tudelft.nl

Funding information

NIH, Grant/Award Number: R01HL153146, R21EB028369 and P41EB027061

Abstract

Purpose: To implement a free-breathing sequence for simultaneous quantification of T_1 , T_2 , and T_2^* for comprehensive tissue characterization of the myocardium in a single scan using a multi-gradient-echo readout with saturation and T_2 preparation pulses.

Methods: In the proposed Saturation And T_2 -prepared Relaxometry with Navigator-gating (SATURN) technique, a series of multi-gradient-echo (GRE) images with different magnetization preparations was acquired during free breathing. A total of 35 images were acquired in 26.5 ± 14.9 seconds using multiple saturation times and T_2 preparation durations and with imaging at 5 echo times. Bloch simulations and phantom experiments were used to validate a 5-parameter fit model for accurate relaxometry. Free-breathing simultaneous T_1 , T_2 , and T_2^* measurements were performed in 10 healthy volunteers and 2 patients using SATURN at 3T and quantitatively compared to conventional single-parameter methods such as SASHA for T_1 , T_2 -prepared bSSFP, and multi-GRE for T_2^* .

Results: Simulations confirmed accurate fitting with the 5-parameter model. Phantom measurements showed good agreement with the reference methods in the relevant range for in vivo measurements. Compared to single-parameter methods comparable accuracy was achieved. SATURN produced in vivo parameter maps that were visually comparable to single-parameter methods. No significant difference between T_1 , T_2 , and T_2^* times acquired with SATURN and single-parameter methods was shown in quantitative measurements (SATURN $T_1 = 1573 \pm 86$ ms, $T_2 = 33.2 \pm 3.6$ ms, $T_2^* = 25.3 \pm 6.1$ ms; conventional methods: $T_1 = 1544 \pm 107$ ms, $T_2 = 33.2 \pm 3.6$ ms, $T_2^* = 23.8 \pm 5.5$ ms; $P > .2$)

Conclusion: SATURN enables simultaneous quantification of T_1 , T_2 , and T_2^* in the myocardium for comprehensive tissue characterization with co-registered maps, in a single scan with good agreement to single-parameter methods.

This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2021 The Authors. *Magnetic Resonance in Medicine* published by Wiley Periodicals LLC on behalf of International Society for Magnetic Resonance in Medicine.

KEYWORDS

cardiac quantitative imaging, free-breathing, T_1 mapping, T_2 mapping, T_2^* mapping

1 | INTRODUCTION

Quantitative mapping in the myocardium has received major clinical interest, as markers related to myocardial relaxation time yield promising sensitivity to a broad spectrum of cardiomyopathies. T_1 , T_2 , and T_2^* mapping are routinely used in advanced CMR centers and received increasing interest in community recommendations and consensus statements for the assessment of ischemia, fibrosis, edema, and amyloidosis or iron deposition.¹⁻⁴

A wide variety of mapping sequences was proposed in the last decades for noninvasively studying the myocardial tissue state.⁵⁻⁹ Myocardial T_1 mapping is most commonly performed based on a series of inversion or saturation recovery images and has shown promise for the assessment of ischemic and nonischemic cardiomyopathies.^{1,4,10,11} While inversion recovery-based methods have shown improved precision and map quality, saturation recovery methods yield more accurate T_1 maps insensitive to the heart rate, the magnetization evolution, and other confounders.¹²⁻¹⁴

In addition to T_1 mapping, myocardial T_2 mapping is increasingly used for the reliable assessment of myocardial edema.¹⁵ State of the art cardiac T_2 mapping is performed by acquiring at least 3 T_2 -prepared balanced steady-state free precession (bSSFP) images to provide robust and reproducible T_2 maps.¹⁵⁻¹⁸

Myocardial T_2^* quantification has demonstrated high clinical value for the assessment of myocardial iron accumulation.¹⁹⁻²¹ According to relevant guidelines, T_2^* measurements in the myocardium is most commonly performed by acquiring 8 echoes with a multi-gradient-echo readout and performing an exponential fit.¹⁹

The methods described above each require one breath-hold per slice. Therefore, free-breathing methods and simultaneous quantification of T_1 and T_2 were proposed to improve patient comfort and shorten measurement time.²²⁻³⁰ Simultaneous T_1 and T_2 mapping was obtained in a single breath-hold by combining saturation/inversion pulses and T_2 preparation modules to improve the detection of abnormalities by inherently co-registered parametric maps.^{22,31,32} This method was expanded to a navigator gated free-breathing approach allowing the coverage of T_1 and T_2 in the entire myocardium in a single scan avoiding deviations due to incorrect breath-holds.^{23,33} Magnetic resonance fingerprinting was proposed for joint estimation of T_1 and T_2 based on undersampled non-Cartesian readouts with varying preparations.²⁵ Most recently, cardiac multitasking was introduced, as a novel method for multiparameter mapping, where contrast and physiological variations

are modeled by a low-dimensional representation, enabling a continuous acquisition of multiparametric 3D maps.²⁴


However, the lack of a combined method for assessment of all 3 clinically relevant tissue characteristics (T_1 , T_2 , and T_2^*) requires multiple sequences in clinical practice, expanding the scan protocol and prolonging examination duration. Furthermore, many recently developed methods rely on implicit or explicit model-based regularization.^{34,35} This often induces quantification inaccuracies and renders the methods' quantification susceptible to changes in the reconstruction pipeline.

In this study, we sought to provide a method for free-breathing assessment of all clinically relevant relaxation times - T_1 , T_2 , and T_2^* . A navigator gated sequence with multi-gradient-echo readout and saturation and T_2 preparation pulses is developed. The accuracy of the proposed technique is evaluated in phantom measurements and in vivo image quality is assessed in healthy subjects and a small cohort of patients.

2 | METHODS

2.1 | Sequence design

Figure 1 depicts the sequence diagram of the proposed Saturation And T_2 prepared Relaxometry with Navigator-gating (SATURN) sequence. The sequence is based on a single-shot multi-gradient-echo readout generating 5 echoes for each end-diastolic imaging window. We used a prospective navigator on the diaphragm of the liver with a gating window of 4-5 mm depending on the subject's breathing pattern. Navigator gating is performed with the following accept-reject scheme: The first contrast without preparation was repeated if the navigator was rejected. Saturation prepared images were also immediately re-attempted in the next heartbeat. No navigator was played during the rest periods before the T_2 preparation. For T_2 -prepared images, T_2 preparation was only performed if the navigator was accepted. In this way, if the navigator was rejected the T_2 -prepared image could be re-attempted immediately, without the need of additional rest-periods. However, in this way, navigator rejections lead to an increase in effective rest periods. We used saturation and T_2 preparation pulses before the readouts to generate T_1 and T_2 contrasts. Therefore, we combined the SASHA 3-parameter fit model with the T_2 -prepared bSSFP 3-parameter fit model. Since we only use short echo times (TE) for the gradient-echo readout and the noise floor for the T_2^* decay is not corrected, we used a truncation model for



$$S(T_S, T_2^p, TE, A, B) = \left(\underbrace{A \left(1 - \exp\left(-\frac{T_S}{T_1}\right)\right)}_{T_1 \text{ Fit}} \cdot \underbrace{\exp\left(-\frac{T_2^p}{T_2}\right) + B}_{T_2 \text{ Fit}} \right) \cdot \underbrace{\exp\left(-\frac{TE}{T_2^*}\right)}_{T_2^* \text{ Fit}}$$

FIGURE 1 Sequence diagram for the proposed T_1 , T_2 , and T_2^* mapping technique. Navigator pulses (light blue) are played before the readouts and the preparation pulses. Five different multi-gradient-echoes per imaging block are generated. The first contrast is performed without any preparation pulses to image the fully relaxed magnetization signal. Second and third contrasts are prepared with 25 and 50 ms T_2 preparation pulses comprising composite hard pulses,³⁷ respectively. A non-selective saturation recovery (WET) pulse is performed immediately (T_S^{\min}) before the readout of contrasts number 4 and 6. The same preparation pulse is played in the systole for contrasts 5 and 7, facilitating longer T_1 relaxation³⁸

T_2^* as previously suggested.³⁶ The 5-parameter truncation fit model is given as

$$S(T_S, T_2^p, TE, A, B) = \left(\underbrace{A \left(1 - \exp\left(-\frac{T_S}{T_1}\right)\right)}_{\text{SASHA Fit}} \cdot \underbrace{\exp\left(-\frac{T_2^p}{T_2}\right) + B}_{T_2 \text{ Fit}} \right) \cdot \underbrace{\exp\left(-\frac{TE}{T_2^*}\right)}_{T_2^* \text{ Fit}} \quad (1)$$

Here, the fitting parameter B is used to account for the T_1 offset. Thus, T_2^* is reconstructed with a truncation model. The first contrast is performed without any preparation representing full magnetization recovery (infinite saturation time, T_S) and T_2 preparation time of $T_2^p = 0$. The second block consists of 2 different T_2 -weighted contrasts using preparation durations of 25 and 50 ms, respectively, as previously recommended.²² Four seconds of rest period were inserted before each image without saturation preparation to allow for full magnetization recovery. Due to the rest-periods, full magnetization recovery was assumed prior to the T_2 preparation. The third block acquires images with saturation preparation to sample the T_1 recovery

curve. The fourth and sixth image is performed with a saturation pulse before the readout to mimic the effect of a very long T_2 preparation³⁷ and short saturation times and, thus, T_S and T_2^p was set to T_S^{\min} and 0. Image 5 and 7 are acquired with saturation preparation with a maximum T_S for maximum precision.³⁸

The full acquisition comprises 7 different contrast preparations followed by imaging at 5 echo times, yielding a total of 35 images. Saturation pulses were performed using a composite “Water suppression Enhanced through T_1 -effects” (WET) pulse to reduce the sensitivity to B_1 .³⁹ The T_2 preparation module consist of a 90° rectangular flip-down pulse, a 270° rectangular flip-up pulse and composite 180° MLEV refocusing pulses in between.^{40,41} Centric k-space reordering was used for increased signal-to-noise ratio and shorter minimum saturation times.

2.2 | Sequence parameters

All measurements were performed on a 3T MRI scanner (Magnetom Skyra, Siemens Healthineers, Erlangen,

Germany) with a 28-channel receiver coil array. Sequence parameters are listed in Table 1.

SATURN was performed using GRAPPA with acceleration factor $R = 3$. Additionally, GRAPPA with acceleration $R = 4$ was explored for the use in subjects with higher heart rates. SPIRiT⁴² with locally low rank (LLR) reconstruction was used for improved noise-resilience at acceleration $R = 4$, as previously proposed.⁴³⁻⁴⁵

2.3 | Simulations

Bloch simulations were used to calculate the magnetization of the proposed SATURN sequence and validate the accuracy of the quantification. All pulse sequences were simulated with the above listed sequence parameters. The magnetization was simulated with time-steps of 0.1 ms. Imaging and preparation pulses were simulated with corresponding rotation matrices with 100% efficiency. The center of the k-space was chosen to extract the signal magnitude. T_1 (1200-1700 ms), T_2 (20-70 ms), and T_2^* (5-60 ms) were varied and the magnitude was fitted with the proposed 5-parameter fit model given in Equation (1). Four confounding factors were included in the simulations: Rest periods before the T_2 preparation pulses were varied between 1 and 10 seconds. For all other simulations, 10 seconds were used to eliminate insufficient recovery as the primary source of inaccuracy. Image noise was added to the simulations. Rician noise was generated with an SNR between 0 and 30 and a Monte Carlo size of 1000. Different heart rates were simulated between 50 and 140 bpm. Finally, imperfect T_2 preparation was simulated by reducing the flip angle of the flip-down and flip-up pulses.

2.4 | Phantom experiments

Phantom measurements were performed to evaluate the accuracy and precision of the proposed SATURN sequence. Reference measurements for T_1 were performed using an inversion-recovery spin echo sequence with $T_I = 100, 200, 500, 1000, 2000, 5000, 8000$ ms, TE/TR = 12/10 000 ms, and imaging geometry as specified above. T_2 reference scans were performed with a spin echo sequence with TE = 17, 30, 50, 100, 150, 250 ms and otherwise identical imaging parameters to the inversion recovery spin-echo (IR-SE). GRE was performed for T_2^* quantification with 12 contrasts ranging from TE = 2-60 ms, TR = 10 000 ms and 1 k-space line per readout with the same imaging parameters listed above. All measurements were additionally compared with single-parameter methods for myocardial mapping (listed in Table 1): SASHA T_1 ⁴⁶ with a minimum and maximum saturation time of 103 ms and 600 ms, T_2 -prepared bSSFP using 4 different T_2 weightings (0 ms, 25 ms, 50 ms, and ∞ ms) and a 3-parameter fit model,^{22,47} and multi-GRE T_2^* with 8 echoes ranging from 1.6 to 16.3 ms¹⁹ using the 2-parameter truncation model.³⁶ The cardiac cycle was simulated and set to a heart rate (HR) of 60 bpm.

2.5 | In vivo experiments

In vivo measurements were performed in 10 healthy volunteers (23-29 years old, 26.1 ± 1.5 years, heart rate: 67.2 ± 7.7 bpm, 3 female), 1 patient (69 years old, female, heart rate: 72 bpm) with hypertrophic cardiomyopathy (HCM), and 1 patient (66 years old, male, heart rate: 79 bpm) with suspected hypertensive heart

Parameters	SATURN	SASHA	T_2 bSSFP	multi-GRE
FOV	384 × 288 mm ²			
In-plane res.	2 × 2 mm ²			
Slice thickness	8 mm			
Partial Fourier	6/8			
Readout	Multi-GRE	bSSFP	bSSFP	multi-GRE
Flip angle	20°	45°	45°	20°
acq. k-Space lines	36	66	66	11
Bandwidth	1530 Hz/px	1130 Hz/px	1130 Hz/px	965 Hz/px
GRAPPA	$R = 3$ or 4	$R = 2$		
Respiration	Free-breathing	Breath-hold (exhaled)		
Number of echoes	5	1	1	8
TE	1.0-8.5 ms	1.3 ms	1.3 ms	1.6-16.3 ms
TR	10.3 ms	2.7 ms	2.7 ms	18.1 ms
Nom. acquisition time	18.5 s	10 s	10 s	8 s

TABLE 1 Sequence parameters for SATURN and the reference methods (SASHA, T_2 -prepared bSSFP, multi-GRE)

Note: Common parameters are depicted with blue shading. Nominal acquisition time is calculated for a heart rate of 60 bpm and a gating efficiency of 50%.

disease (HHD) after written consent was obtained. All images were acquired in the mid-ventricular short-axis view using the parameters described in the previous section.

SATURN was performed with a maximum T_S adjusted to the subject's heart rate. Motion between images from different heartbeats was reduced by retrospective image registration. Rigid registration was performed with mutual information in the region of interest as the similarity metric. Voxel-wise fitting was performed using the 5-parameter model.

Regions of interest were manually drawn in the entire myocardium, with careful distancing to the epi- and endocardial borders. Bullseye plots were generated for the 6 mid-ventricular segments of the American Heart Association (AHA) segment model.⁴⁸

Standard deviation maps (SD maps) were generated by calculating all partial derivatives of the fit function as previously proposed.⁴⁹ The covariance matrix is calculated by the inverse of the Hessian matrix. The square root of the sum of the diagonal entries of the covariance matrix is used as an approximation for the voxel-wise SD of the individual parameters.

2.6 | Statistics

The within-segment mean and the within-segment SD of the T_1 , T_2 , and T_2^* times were averaged across all subjects.

Additionally, the within-segment means of the SD T_1 , T_2 , and T_2^* times were calculated using the corresponding voxel-wise SD maps. Intersubject variability was calculated as the SD of the within-segment mean across all subjects. Pair-wise comparison was performed using Student's t-tests using the Bonferroni correction for multiple comparisons along T_1 , T_2 , and T_2^* . Values of p less than 0.05 were considered significant. Significance between segments of the myocardium was tested using the ANOVA test. Relative deviations were compared by dividing the absolute difference between reference and SATURN with the reference.

3 | RESULTS

3.1 | Simulations

Figure 2A shows the simulated longitudinal magnetization evolution of the proposed SATURN sequence with varying T_1 , T_2 , and T_2^* . Figure 2B plots the fitted relaxation times against the reference relaxation times to depict the measurement accuracy. Accurate multiparameter quantification for T_1 , T_2 , and T_2^* across the relevant in vivo range ($T_1 = 800\text{--}2200$ ms, $T_2 = 30\text{--}70$ ms, $T_2^* = 10\text{--}60$ ms) was achieved in simulations. One source of deviation for T_2 was incomplete recovery during the rest-periods leading to very slight deviations




FIGURE 2 A, Simulated magnetization evaluation of the proposed sequence for varying T_1 (800–2000 ms), T_2 (30–100 ms), and T_2^* (20–100 ms) on the top. Increasing relaxation times are depicted by increasing brightness. B, Bottom panel shows the proposed 5-parameter fit (blue) to the used relaxation time

in T_2 (0.02% for 50 ms, <5% deviation for 100 ms) as shown in Supporting Information Figure S1. T_2^* quantification was found to be more susceptible to higher noise levels than T_1 and T_2 . T_1 , T_2 , and T_2^* accuracy were independent of the heart rate. T_2 accuracy was additionally compromised by an imperfect T_2 preparation efficiency resulting in a strong underestimation, especially for longer T_2 times.

3.2 | Phantom

Phantom measurements (Figure 3A) showed good agreement with reference methods. Deviations of less than 7.7% for relaxation times across the relevant in vivo range were

observed. In Figure 3B, the relative difference of the measured relaxation times to the reference is shown as well as exemplary maps for SATURN and the reference are shown below (Figure 3C). SATURN T_1 times compared with the inversion recovery spin-echo, yielding accuracy comparable to SASHA. T_2 times were accurate in the relevant range (5.2% deviation) and decreased when exceeding 100 ms with relative deviations of up to 20%. For T_2^* of less than 100 ms T_2^* accuracy (7.7% deviation) was slightly higher compared with the conventional single-parameter method, where a decrease of up to 11 ms was measured compared with the reference GRE. SATURN overestimates long T_2^* times compared with the GRE and multi-GRE.⁵⁰ All representative relaxation times per tube are displayed in Supporting Information Table S1.




FIGURE 3 A, T_1 , T_2 , and T_2^* acquired with SATURN (blue) and the single-parameter methods (red) in a phantom plotted against the reference values. For T_2^* , 2 tubes with high relaxation times are outside of the depicted range. B, Relative difference between the reference method and SATURN and the single-parameter models for the different relaxation times. C, Representative T_1 , T_2 , and T_2^* maps for SATURN and the reference

3.3 | In vivo

The average acquisition time for SATURN in the 10 healthy subjects was 26.5 ± 14.9 seconds, which corresponds to an average gating efficiency of $54\% \pm 30\%$. The minimal T_S^{\min} was 7 ms for every subject and the maximal T_S^{\max} was 601 ± 65 ms. An example of magnitude data acquired with SATURN in 1 healthy subject is shown in Figure 4A. Signal intensities from the septum are plotted across 35 measurements along with the fitted signal model (Figure 4B). Visual image quality is high for T_1 and T_2 . Artifacts are observed in T_2^* maps (Figure 5A). SD maps depict the homogeneous mapping precision throughout the myocardium (Figure 5B).

Examples of quantitative parameter maps acquired with SATURN compared with the single-parameter reference methods are shown in Figure 5 for 1 healthy subjects (2 more subjects are shown in Supporting Information Figure S2). Visual image quality is comparable with the single-parameter scans for T_1 and T_2 . However, some blurring is observed in the SATURN maps. T_1 and T_2 maps depict a homogeneous myocardium clear of artifacts. T_2^* maps acquired with SATURN appear visually smoother than the reference.

Figure 6 shows the in vivo mean T_1 , T_2 , and T_2^* times for SATURN over the conventional methods for all healthy subjects. Below the Bland-Altman plot is depicted. A bias of +29.16 ms was measured for T_1 and a bias of +1.54 ms was measured for T_2^* . T_2 times yielded negligible bias compared with T_1 and T_2^* but limits of agreement of ± 9.4 ms. All representative relaxation times per subject are displayed in Supporting Information Table S2.

Supporting Information Figure S3 shows the difference between SATURN acquired with GRAPPA with acceleration

factor $R = 3$, $R = 4$, and $R = 4$ using SPIRiT + LLR regularization. T_2 map quality shows only minor differences between $R = 3$ or $R = 4$ with deviations of less than 2%. However, T_1 map quality is improved with 36.2% lower within-segment SDs for $R = 3$ compared with $R = 4$. Precision is regained by using regularization (SPIRiT + LLR) and image quality is visually improved (only 5.4% lower within-segment SDs). SATURN T_1 maps appear smoother and more homogeneous when using $R = 3$ with smaller variations within the myocardium. Additional artifacts appear in T_2^* maps using $R = 4$, which are largely alleviated using regularization.

Figure 7 represents the AHA 6 segment bullseye plots showing the mean quantitative measures across all healthy for the T_1 , T_2 , and T_2^* and the corresponding within-segment SD. The relaxation times in the healthy myocardium measured with SATURN averaged over all 6 AHA segments were $T_1 = 1573 \pm 86$ ms, $T_2 = 33.2 \pm 3.6$ ms, comparable to the conventional methods ($T_1 = 1544 \pm 107$ ms; $P = .22$, $T_2 = 33.2 \pm 3.6$ ms; $P = .98$). T_2^* obtained with SATURN was 25.3 ± 6.1 ms, corresponding to a 5.9% increase compared to the conventional method (23.8 ± 5.3 ms; $P = .33$) with both methods suffering from artifacts. No significant differences were found between the in vivo times measured with SATURN and the conventional methods for neither T_1 , T_2 or T_2^* .

No significant differences among segments were measured for SATURN T_1 ($P = .36$) but significant differences for T_2 ($P = .037$) and T_2^* ($P = .038$), with the lowest T_2/T_2^* times being observed in the mid-inferior segment. The same trend is observed for the conventional methods. For SASHA T_1 , no significant difference among the segments ($P = .83$) was observed, but significant differences for the single-parameter T_2 ($P = .033$) and T_2^* ($P < .01$), depicting




FIGURE 4 A, Magnitude images from the septum are plotted across the 35 measurements. B, Image intensities of the acquired (blue) and fitted (orange) signal model and the fit residual of a voxel in the septal myocardium are shown below where the gray area marks deviations of less than 1 standard deviation




FIGURE 5 A, In vivo T_1 , T_2 , and T_2^* maps acquired with the proposed SATURN sequence (left) and single-parameter reference methods (right) for 1 healthy subject. Visually homogeneous mapping is achieved throughout the myocardium for T_1 and T_2 , minor artifacts appear in T_2^* maps. Image quality appears visually comparable to the reference methods. B, Below the standard deviation (SD) maps are shown for the 3 relaxation times and the same subject for SATURN and the reference methods

a similar drop in the mid-inferior segment. Intersubject variability of 57.9 ms (3.7% compared with the mean value) was observed in T_1 , which is higher than for SASHA (42.3 ms (2.7%)). Intersubject variability of 3.3 ms (9.9%) for T_2 obtained with SATURN were in the range of the T_2 -prepared bSSFP with 3.2 ms (9.6%), and 3.6 ms (14.2%) for T_2^* compared with the multi-GRE 3.2 ms (13.4%) were observed.

SD maps are calculated for all healthy subjects for SATURN and the conventional methods and resulted in mean values of $\sigma(T_1) = 68$ ms, $\sigma(T_2) = 1.1$ ms and $\sigma(T_2^*) = 3.3$ ms and for the conventional methods $\sigma(T_1) = 39.3$ ms, $\sigma(T_2) = 1.9$ ms and $\sigma(T_2^*) = 1.5$ ms. Examples of SD maps are shown in Figure 5B and Supporting Information Figure S2. Figure 8 shows the mean and the SD of the calculated SD maps in each of the 6 segments. For T_1 , SATURN achieved 23.3% lower within-segment SDs and improved precision compared with SASHA T_1 map. T_2 shows comparable precision between SATURN and the single-parameter method (5.1% deviations). Increased within-segment SDs of 8.3% are observed for SATURN T_2^* compared with the reference multi-GRE.

Figure 9 shows SATURN T_1 , T_2 , and T_2^* maps for a patient with HCM, and 1 patient with suspected HHD and the corresponding bullseye plots. Increased T_1 , T_2 , and T_2^* times (1607/47.0/35.5 ms vs 1487/38.5/26.5 ms) are observed in the septal regions compared with the lateral myocardium in the patient with HCM. SATURN shows increased T_1 times

and patchy structures in the patient with HHD. T_2^* times are substantially elevated.

4 | DISCUSSION

In this study, we proposed the SATURN sequence for free-breathing simultaneous quantification of T_1 , T_2 , and T_2^* in the myocardium based on a gradient-echo readout in combination with saturation pulses and T_2 preparation pulses. We demonstrated good agreement with Bloch simulations and phantom experiments yielding generally accurate T_1 times. However some biases for T_2 and T_2^* are observed. In vivo measurements provided robust image quality comparable to reference methods for all segments in the mid-ventricular short-axis view.

T_1 measurements resulted in good accuracy compared to spin-echo sequences and SASHA in the phantom and in vivo. Mean T_1 times in the 6 segments are comparable to previously reported values for saturation based T_1 mapping at 3T.^{12,50} T_1 maps yielded similar image quality and smaller within-segment standard deviations compared with SASHA. Similar inter-subject variability was found between SATURN and the reference method.

Our simulations indicate that accuracy in T_2 mapping is compromised for long T_1/T_2 combinations due to insufficient recovery during the rest-period. However, as this effect is only marked at values outside the relevant in vivo range, a




FIGURE 6 A, SATURN over the reference T_1 , T_2 , and T_2^* times for each healthy subject. The black line shows the bisector. B, Bland-Altman plot showing the difference between SATURN and the reference over the mean of both. The legend shows the p-value of the Student's t-test




FIGURE 7 Bullseye plot of T_1 , T_2 , and T_2^* relaxation times acquired with SATURN and the single-parameter reference. The bullseye values are reported as the within-segment mean \pm the within-segment standard deviation average across all healthy subjects. Small differences between SATURN and the reference was observed for T_1 and T_2 . T_2^* obtained with SATURN was 5.9% increased compared with the reference




FIGURE 8 Bullseye plot of T_1 , T_2 , and T_2^* relaxation times acquired with SATURN and the single-parameter reference. The bullseye values are reported as the within-segment mean \pm the within-segment standard deviation of the standard deviation map (SD map) for each segment. The voxel-wise standard deviation was higher for T_1 and T_2^* obtained with SATURN and smaller for T_2 compared with the single-parameter methods

choice of 4 seconds rest periods seemed justified. In vivo T_2 measurements resulted in lower T_2 times than reported in literature,^{22,51,52} but only minor differences were observed between SATURN and the reference T_2 -prepared bSSFP. The lower T_2 times obtained in this study as compared to previous literature^{22,51,52} are related to the use of a three-parameter fit model, which was previously shown to yield lower T_2 times (Supporting Information Figure S4). Other than that, we observed a drop of T_2 in mid-inferior segments in some healthy subjects due to B_1^+ inhomogeneities, which could be corrected by better shimming routines. We used rest periods before the T_2 preparations instead of saturation pulses directly after the ECG trigger because the SNR of the T_2 -prepared images for the gradient-echo readout was too low for accurate T_2 quantification as numerical simulations showed.⁵³ T_2 maps in vivo and in phantom appeared visually smoother and more blurred as compared with the conventional single-parameter maps due to the centric k-space reordering in SATURN. With centric k-space reordering, the magnetization transfer function acts as a slight low pass filter.⁵⁴ We decided in favor of centric k-space reordering due to the improved quantification result and image quality, especially for T_1 .

Bloch simulations without noise result in accurate T_2 quantification. However, phantom measurements resulted in deviations of up to 20%, likely due to susceptibility artifacts and increased noise, as this was the dominant factor in the Bloch simulations. Especially for the tubes with very high T_2^* times the quantification in the phantom failed, which might be due to the very short maximum TE of the 5 echoes from SATURN. However, for T_2^* in the in vivo relevant range

SATURN was still observed to be more accurate than the reference GRE method. T_2^* times in vivo are in the range of reported literature⁵⁵⁻⁵⁷ and slightly increased compared with the reference GRE (5.9%). The overestimation is likely linked to a shorter maximum TE. However, increased T_2^* times are measured without truncation due to sufficient SNR.^{36,57} A drop in T_2^* was observed in the mid-inferior segment due to B_1^+ inhomogeneities as also observed for T_2 .

Higher accelerations might be necessary for patients with high heart rates to reduce the time per single-shot acquisition. The variability in T_1 maps is increased when using higher acceleration factors (GRAPPA $R = 4$). However, this can be alleviated by using regularization (SPIRiT + LLR) at the cost of inducing complexity in the post-processing. T_2 maps reconstructed using acceleration factors of $R = 3$ and $R = 4$ resulted in visually similar T_2 maps with only slight deviations of 2.1% in the T_2 times and 1.9% in the within-segment standard deviations. Similar to T_1 , for the T_2^* the use of $R = 4$ increases the within-segment standard deviation by 52.4%, which might be due to the low SNR for images with long TEs.

In the patient with HCM we observed an increased T_1 , T_2 , and T_2^* time as reported in literature.⁵⁸⁻⁶⁰ Image quality was visually good. For the patient with suspected HHD, increased T_1 and T_2^* was observed in the septal region and patchy structures in the T_1 map as typically observed in HHD.⁶¹ No reference methods were acquired in patients, which will be evaluated in future work.

Simultaneous measurements of T_1 , T_2 , and T_2^* is more time-efficient since all parameters are acquired in 1 scan (average




FIGURE 9 SATURN T_1 , T_2 , and T_2^* maps for a patient with hypertrophic cardiomyopathy (HCM) and 1 patient with suspected hypertensive heart disease (HHD). The corresponding bullseye plots are shown respectively. In the patient with HCM, increased T_1 , T_2 , and T_2^* were observed in the septal region. For the patient with HHD, increased T_1 and T_2^* were observed in the septal region as well as patchy structures in the T_1 map

acquisition time was 26.5 ± 14.9 seconds). Additionally, they share the same volume and are, therefore, inherently co-registered. This eases the fusion of imaging information as corresponding regions are easy to identify. Furthermore, the assessment of multiple quantitative measures increases the specificity for diagnosis.^{1,2,51}

Free-breathing imaging was achieved by using a prospective navigator on the liver diaphragm. This may minimize the susceptibility to incomplete breath-holds as often observed in patients suffering from dyspnea. Residual motion is compensated by the use of image registration. We used rigid-registration as previously reported to yield satisfactory results in healthy subjects (Supporting Information Figure S5).⁶² In patients with variable breathing patterns and/or arrhythmia, the motion correction for respiratory as well as the cardiac cycle might be improved by using non-rigid registrations,

which is subject of future work. In addition, simultaneous multislice acquisition⁶³ can be used to cover multiple slices per acquisition, which enables whole heart imaging in a relatively short time.

Intramyocardial fat is often present in cardiac patients and is known to shorten the T_1 and T_2 times.⁶⁴ While variable impact of the fat fraction on bSSFP based cardiac relaxometry has been reported,⁶⁵ the effects on GRE-based mapping, as proposed in this study, are expected to be affected by fewer confounders. Furthermore, in the presence of substantial intramyocardial fat, the T_2^* decay deviates from a monoexponential decay. Dixon-encoding mapping might be used to separate the fat and water signal and overcome the deviations in the quantitative measures.^{66,67} Integration of these techniques in our proposed sequence and dedicated evaluation for fatty storage disease warrant further investigation.

A physics-based 5-parameter model was used for the quantification. Recent trends emerged using machine learning for improving the reconstruction and fitting with non-explicit modeling and might be applied due to the limited spatial resolution, partial volume effects, and noise.⁶⁸⁻⁷¹

This study has several limitations. Saturation recovery based methods for T_1 quantification suffer from a decreased dynamic range of the T_1 recovery curve, which is known to decrease the precision.³⁸ However, compared with inversion recovery methods such as MOLLI, the accuracy is not impacted⁷² (Supporting Information Figure S1). The dynamical range could be increased by shifting the readout to the succeeding heart-beat, as previously reported.⁷³ However, in this case, navigator gating may affect the sampling of the saturation recovery. Nonetheless, this modification may lead to valuable improvements in terms of map quality for tachycardiac patients and warrants further investigation. Single-shot imaging suffers from long readout blocks, especially for a multi-gradient-echo readout with 5 echoes. Higher heart rates will result in more cardiac motion during the acquisition. Therefore, the maximal TR of the echoes has to be short enough to acquire the whole k-space in 1 diastolic phase. However, short TR reduces the accuracy of the T_2^* quantification of long T_2^* times as observed under certain circumstances or lower field-strength. Higher acceleration factors enable the sampling of longer echo times in the same acquisition window, albeit at the cost of reduced SNR. We showed that this limitation might be partially compensated for by the use of regularization when using acceleration factors higher than $R = 3$. A maximum TE of 8.6 ms is short compared with conventional methods that often use a maximum TE around 16-18 ms.²⁰ We decided to use a truncation fitting model to increase the quantification accuracy, especially for the low SNR contrasts 4 and 6.³⁶ Nevertheless, the use of short echo times might lead to an overestimation of T_2^* . However, an increase of 1.5 ms in T_2^* in vivo compared with the conventional multi-GRE was obtained with SATURN. This deviation is explained by the shorter maximum TE relative to the reference method. Increasing the length of the GRE readout train may be considered in a trade-off against higher acceleration rates if improved accuracy for long T_2^* is desired. Faster acquisition schemes such as radial single-shot images might offer a better compromise between longer TE and short enough acquisition windows, which will be evaluated in further research. Conventionally, T_2^* maps are acquired with lower spatial resolution compared with T_1 and T_2 . Since we are measuring all 3 parameters from the same scan with the same spatial resolution we acquire with slightly higher resolution for T_2^* as commonly acquired.²⁰ Furthermore, it is generally recommended to perform T_2^* mapping at 1.5T. Hence, the quality of the T_2^* quantification might show superior results at 1.5T. Blood signal suppression is also often used in T_2^* mapping to alleviate partial volume effects. However, in this study, we

refrained from additional blood signal suppression but may benefit from decreased partial voluming due to an increased imaging resolution.

5 | CONCLUSION

SATURN enables joint quantification of the most relevant clinical relaxation times, T_1 , T_2 , and T_2^* , with robust image quality in a single free-breathing scan. Good quantification accuracy was demonstrated in a phantom. In vivo free-breathing imaging yielded high visual image quality.

ACKNOWLEDGEMENTS

The authors acknowledge grant support by the 4TU Federation, ZonMW OffRoad 04510011910073, a NWO Startup STU.019.024 and the NIH R01HL153146, R21EB028369, P41EB027061.

ORCID

Ingo Hermann <http://orcid.org/0000-0002-6379-5299>

REFERENCES

- Messroghli DR, Moon JC, Ferreira VM, et al. Clinical recommendations for cardiovascular magnetic resonance mapping of T_1 , T_2 , T_2^* and extracellular volume: a consensus statement by the society for cardiovascular magnetic resonance (SCMR) endorsed by the European association for cardiovascular imaging (EACVI). *J Cardiovasc Magn Reson*. 2017;19:75.
- Lota AS, Gatehouse PD, Mohiaddin RH. T_2 mapping and T_2^* imaging in heart failure. *Heart Fail Rev*. 2017;22:431-440.
- Dall'Armellina E, Piechnik SK, Ferreira VM, Si Quang L. Cardiovascular magnetic resonance by non contrast T_1 -mapping allows assessment of severity of injury in acute myocardial infarction. *J Cardiovasc Magn Reson*. 2012;14:15.
- Radenkovic D, Weingärtner S, Ricketts L, Moon JC, Captur G. T_1 mapping in cardiac MRI. *Heart Failure Rev*. 2017;22:415-430.
- Ferreira VM, Piechnik SK, Robson MD, Neubauer S, Karamitsos TD. Myocardial tissue characterization by magnetic resonance imaging: novel applications of T_1 and T_2 mapping. *J Thoracic Imaging*. 2014;29:147-154.
- Hamlin SA, Henry TS, Little BP, Lerakis S, Stillman Arthur E. Mapping the future of cardiac MR imaging: case-based review of T_1 and T_2 mapping techniques. *RadioGraphics*. 2014;34:1594-1611.
- Moon JC, Messroghli DR, Kellman P, et al. Myocardial T_1 mapping and extracellular volume quantification: a society for cardiovascular magnetic resonance (SCMR) and CMR working group of the European Society of Cardiology consensus statement. *J Cardiovasc Magn Reson*. 2013;15:92.
- Messroghli DR, Radjenovic A, Kozerke S, Higgins DM, Sivanathan MU, Ridgway JP. Modified Look-Locker inversion recovery (MOLLI) for high-resolution T_1 mapping of the heart. *Magn Reson Med*. 2004;52:141-146.
- Weingärtner S, Akçakaya M, Basha T, et al. Combined saturation/inversion recovery sequences for improved evaluation of scar and diffuse fibrosis in patients with arrhythmia or heart rate variability. *Magn Reson Med*. 2014;71:1024-1034.

10. Haaf P, Garg P, Messroghli DR, Broadbent DA, Greenwood JP, Plein S. Cardiac T_1 Mapping and Extracellular Volume (ECV) in clinical practice: a comprehensive review. *J Cardiovasc Magn Reson*. 2016;18:89.
11. Puntmann VO, Peker E, Chandrasekhar Y, Nagel E. T_1 Mapping in characterizing myocardial disease a comprehensive review. *Circulation Res*. 2016;119:277-299.
12. Weingärtner S, Meßner NM, Budjan J, et al. Myocardial T_1 -mapping at 3T using saturation-recovery: reference values, precision and comparison with MOLLI. *J Cardiovasc Magn Reson*. 2017;18:84.
13. Kellman P, Yue H, Chow K, Spottiswoode BS, Arai AE, Thompson RB. Optimized saturation recovery protocols for T_1 -mapping in the heart: influence of sampling strategies on precision. *J Cardiovasc Magn Reson*. 2014;16:55.
14. Hermann I, Uhrig T, Chacon-Caldera J, Akcakaya M, Schad LR, Weingärtner S. Towards measuring the effect of flow in blood T_1 assessed in a flow phantom and in vivo. *Phys Med Biol*. 2020;65.
15. Verhaert D, Thavendiranathan P, Giri S, et al. Direct T_2 Quantification of Myocardial Edema in Acute Ischemic Injury. *JACC*. 2011;4:269-278.
16. Kellman P, Aletras AH, Mancini C, McVeigh ER, Arai AE. T_2 -prepared SSFP improves diagnostic confidence in edema imaging in acute myocardial infarction compared to turbo spin echo. *Magn Reson Med*. 2007;57:891-897.
17. Giri S, Chung Y-C, Mancini A, Rajagopalan S, Raman SV, Simonetti OP. T_2 quantification for improved detection of myocardial edema. *J Cardiovasc Magn Reson*. 2009;11:56.
18. Basha T, Akcakaya M, Roujol S, Nezafat R. Precision and reproducibility of T_2 quantifications in myocardial T_2 mapping: impact of the number of echoes and reconstruction model. *J Cardiovasc Magn Reson*. 2015;17.
19. Westwood M, Anderson LJ, Firmin DN, et al. A single breath-hold multiecho T_2^* cardiovascular magnetic resonance technique for diagnosis of myocardial iron overload. *J Magn Reson Imaging*. 2003;18:33-39.
20. Kellman P, Xue H, Spottiswoode BS, et al. Free-breathing T_2^* mapping using respiratory motion corrected averaging. *J Cardiovasc Magn Reson*. 2015;3:1741-1753.
21. Triadyaksa P, Oudkerk M, Sijens PE. Cardiac T_2^* mapping: techniques and clinical applications. *J Magn Reson Imaging*. 2019;52(5):1340-1351. <https://doi.org/10.1002/jmri.27023>.
22. Akcakaya M, Weingärtner S, Basha TA, Roujol S, Bellm S, Nezafat R. Joint myocardial T_1 and T_2 mapping using a combination of saturation recovery and T_2 -preparation. *Magn Reson Med*. 2016;76:888-896.
23. Qi H, Bustin A, Cruz G, et al. Free-running simultaneous myocardial T_1/T_2 mapping and cine imaging with 3D whole-heart coverage and isotropic spatial resolution. *Magn Reson Imaging*. 2019;63:159-169.
24. Christodoulou AG, Shaw JL, Nguyen C, et al. Magnetic resonance multitasking for motion-resolved quantitative cardiovascular imaging. *Nat Biomed Eng*. 2018;2:215-226.
25. Hamilton JI, Jiang Y, Chen Y, et al. MR fingerprinting for rapid quantification of myocardial T_1 , T_2 , and proton spin density. *Magn Reson Med*. 2017;77:1446-1458.
26. Cruz G, Jaubert O, Botnar RM, Prieto C. Cardiac magnetic resonance fingerprinting: technical developments and initial clinical validation. *Curr Cardiol Rep*. 2019;21:91.
27. Shao J, Zhou Z, Nguyen K-L, Finn JP, Hu P. Accurate, precise, simultaneous myocardial T_1 and T_2 mapping using a radial sequence with inversion recovery and T_2 preparation. *NMR Biomed*. 2019;32:e4165.
28. Blume U, Lockie T, Stehning C, et al. Interleaved T_1 and T_2 relaxation time mapping for cardiac applications. *J Magn Reson Imaging*. 2009;29:480-487.
29. Weingärtner S, Roujol S, Akcakaya M, Basha TA, Nezafat R. Free-breathing multislice native myocardial T_1 mapping using the slice-interleaved T_1 (STONE) sequence. *Magn Reson Med*. 2015;74:115-124.
30. Ding H, Manuel L, Schär M, et al. Three-dimensional whole-heart T_2 mapping at 3T. *Magn Reson Med*. 2015;74:803-816.
31. Santini F, Kawel-Boehm N, Greiser A, Bremerich J, Bieri O. Simultaneous T_1 and T_2 quantification of the myocardium using cardiac balanced-SSFP inversion recovery with interleaved sampling acquisition (CABIRIA). *Magn Reson Med*. 2015;74:365-371.
32. Sofia Kvernby, Bertus Warntjes Marcel Jan, Henrik Haraldsson, Carl-Johan Carlhäll, Jan Engvall, Tino Ebbers. Simultaneous three-dimensional myocardial T_1 and T_2 mapping in one breath hold with 3D-QALAS. *J Cardiovasc Magn Reson*. 2014;16:102.
33. Guo R, Chen Z, Herzka DA, Luo J, Ding H. A three-dimensional free-breathing sequence for simultaneous myocardial T_1 and T_2 mapping. *Magn Reson Med*. 2019;81:1031-1043.
34. Fessler J. Model-Based Image Reconstruction for MRI. *IEEE Signal Process Mag*. 2010;27:81-89.
35. Mohsin YQ, Lingala SG, DiBella E, Jacob M. Accelerated dynamic MRI using patch regularization for implicit motion compensation. *Magn Reson Med*. 2017;77:1238-1248.
36. He T, Gatehouse PD, Smith GC, Mohiaddin RH, Pennell DJ, Firmin DN. Myocardial T measurements in iron-overloaded thalassemia: An in vivo study to investigate optimal methods of quantification. *Magn Reson Med*. 2008;60:1082-1089.
37. Akcakaya M, Basha TA, Weingärtner S, Roujol S, Berg S, Nezafat R. Improved quantitative myocardial T_2 mapping: impact of the fitting model. *Magn Reson Med*. 2015;74:93-105.
38. Akcakaya M, Weingärtner S, Roujol S, Nezafat R. On the selection of sampling points for myocardial T_1 mapping. *Magn Reson Med*. 2015;73:1741-1753.
39. Ogg RJ, Kingsley RB, Taylor JS. WET, a T_1 - and B1-Insensitive Water-Suppression Method for in Vivo Localized 1H NMR Spectroscopy. *J Magn Reson Ser B*. 1994;104:1-10.
40. Brittain JH, Hu BS, Wright GA, Meyer CH, Macovski A, Nishimura DG. Coronary angiography with magnetization-prepared T_2 contrast. *Magn Reson Med*. 1995;33:689-696.
41. Nezafat R, Stuber M, Ouwkerk R, Gharib AM, Desai MY, Pettigrew RI. B1-insensitive T_2 preparation for improved coronary magnetic resonance angiography at 3 T. *Magn Reson Med*. 2006;55:858-864.
42. Lustig M, Pauly JM. SPIRiT: iterative self-consistent parallel imaging reconstruction from arbitrary k-space. *Magn Reson Med*. 2010;64:457-471.
43. Yaman B, Weingärtner S, Kargas N, Sidiropoulos ND, Akcakaya M. Locally Low-Rank tensor regularization for high-resolution quantitative dynamic MRI. *2017 IEEE 7th International Workshop on Computational Advances in Multi-Sensor Adaptive Processing (CAMSAP)*. 2017:1-5.
44. Moeller S, Weingärtner S, Akcakaya M. Multi-scale locally low-rank noise reduction for high-resolution dynamic quantitative

- cardiac MRI. *2017 39th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*. 2017:1473-1476.
45. Zhang T, Pauly JM, Levesque IR. Accelerating parameter mapping with a locally low rank constraint. *Magn Reson Med*. 2015;73:655-661.
 46. Chow K, Flewitt JA, Green JD, Pagano JJ, Friedrich MG, Thompson RB. Saturation recovery single-shot acquisition (SASHA) for myocardial T_1 mapping. *Magn Reson Med*. 2014;71:2082-2095.
 47. Huang T-Y, Liu Y-J, Stemmer A, Poncelet BP. T_2 measurement of the human myocardium using a T_2 -prepared transient-state true-FISP sequence. *Magn Reson Med*. 2007;57:960-966.
 48. Cerqueira MD, Weissman NJ, Dilsizian V, et al. Standardized myocardial segmentation and nomenclature for tomographic imaging of the heart. *Circulation*. 2002;105:539-542.
 49. Kellman P, Arai AE, Xue H. T_1 and extracellular volume mapping in the heart: estimation of error maps and the influence of noise on precision. *J Cardiovasc Magn Reson*. 2013;15:56.
 50. Roy C, Slimani A, Meester C, et al. Age and sex corrected normal reference values of T_1 , $T_2T_2^*$ and ECV in healthy subjects at 3T CMR. *J Cardiovasc Magn Reson*. 2017;19:1741-1753.
 51. Kim PK, Hong YJ, Im DJ, et al. Myocardial T_1 and T_2 Mapping: techniques and clinical applications. *Korean J Radiol*. 2017;18:113-131.
 52. Pons-Llado G. Reference Normal values for myocardial T_1 and T_2 maps with the MAGNETOM vida 3T system and case examples form clinical practice. *MAGNETOM Flash*. 2019;72:29-33.
 53. Heeswijk RB, Piccini D, Feliciano H, Hullin R, Schwitler J, Stuber M. Self-navigated isotropic three-dimensional cardiac T_2 mapping. *Magn Reson Med*. 2015;73:1549-1554.
 54. Qin Q. Point spread functions of the T_2 decay in k-space trajectories with long echo train. *Magn Reson Imaging*. 2012;30:1134-1142.
 55. Kritsaneepaiboon S, Ina N, Chotsampancharoen T, Roymanee S, Cheewatanakornkul S. The relationship between myocardial and hepatic T_2 and T_2^* at 15T and 3T MRI in normal and iron-overloaded patients. *Acta Radiol*. 2018;59:355-362.
 56. Zaman A, Higgins DM, Kouwenhoven M, Kidambi A, Greenwood JP, Plein S. Robust myocardial T_2 and T_2^* mapping at 3T. *JCMR*. 2012;14.
 57. Sandino CM, Kellman P, Arai AE, Hansen MS, Xue H. Myocardial T_2^* mapping: influence of noise on accuracy and precision. *J Cardiovasc Magn Reson*. 2015;17:7.
 58. Xu J, Zhuang B, Sirajuddin A, et al. MRI T_1 Mapping in hypertrophic cardiomyopathy: evaluation in patients without late gadolinium enhancement and hemodynamic obstruction. *Radiology*. 2020;294:275-286.
 59. Ismail MF, Jabbour A, Gulati A, et al. Role of T_1 and T_2 -mapping in assessing the myocardial interstitium in hypertrophic cardiomyopathy: a cardiovascular magnetic resonance study. *J Cardiovasc Magn Reson*. 2013;15:O32. <https://doi.org/10.1186/1532-429X-15-S1-O32>.
 60. Gastl M, Gotschy A, von Spiczak J, et al. Cardiovascular magnetic resonance T_2^* mapping for structural alterations in hypertrophic cardiomyopathy. *Eur J Radiol Open*. 2019;6:78-84.
 61. Reiter U, Reiter C, Kräuter C, Fuchsjäger M, Reiter G. Cardiac magnetic resonance T_1 mapping Part 2: diagnostic potential and applications. *Eur J Radiol*. 2018;109:235-247.
 62. Zhou R, Huang W, Yang Y, et al. Simple motion correction strategy reduces respiratory-induced motion artifacts for k-t accelerated and compressed-sensing cardiovascular magnetic resonance perfusion imaging. *J Cardiovasc Magn Reson*. 2018;20:23.
 63. Weingärtner S, Moeller S, Schmitter S, et al. Simultaneous multislice imaging for native myocardial T_1 mapping: improved spatial coverage in a single breath-hold. *Magn Reson Med*. 2017;78:462-471.
 64. Kellman P, Bandettini WP, Mancini C, Hammer-Hansen S, Hansen MS, Arai AE. Characterization of myocardial T_1 -mapping bias caused by intramyocardial fat in inversion recovery and saturation recovery techniques. *J Cardiovasc Magn Reson*. 2015;17.
 65. Rossi G, Hilbert T, Mackowiak Adele LC, Pierzchała K, Kober T, Bastiaansen J. Fat fraction mapping using bSSFP Signal Profile Asymmetries for Robust multi-Compartment Quantification (SPARCQ). *Med Phys*. 2020.
 66. Milotta G, Bustin A, Jaubert O, Neji R, Prieto C, Botnar RM. 3D whole-heart isotropic-resolution motion-compensated joint T_1/T_2 mapping and water/fat imaging. *Magn Reson Med*. 2020;84:3009-3026.
 67. Jaubert O, Cruz G, Bustin A, et al. Water-fat Dixon cardiac magnetic resonance fingerprinting. *Magn Reson Med*. 2020;83:2107-2123.
 68. Feng L, Ma D, Liu F. Rapid MR relaxometry using deep learning: an overview of current techniques and emerging trends. *NMR Biomed*. 2020;e4416.
 69. Vidyullatha P, Rao D. Machine learning techniques on multidimensional curve fitting data based on R-square and chi-square methods. *Int J Electric Comp Eng*. 2016;6:974.
 70. Lundervold AS, Lundervold A. An overview of deep learning in medical imaging focusing on MRI. *Z Med Phys*. 2019;29:102-127. Special Issue: Deep Learning in Medical Physics.
 71. Maier A, Syben C, Lasser T, Riess C. A gentle introduction to deep learning in medical image processing. *Z Med Phys*. 2019;29:86-101. Special Issue: Deep Learning in Medical Physics.
 72. Weingärtner S, Akçakaya M, Berg S, Kissinger KV, Manning WJ, Nezafat R. Heart-rate independent myocardial T_1 -mapping using combined saturation and inversion preparation pulses. *J Cardiovasc Magn Reson*. 2013;15:P46.
 73. Meßner NM, Budjan J, Loßnitzer D, et al. Saturation-recovery myocardial T_1 -mapping during systole: accurate and robust quantification in the presence of arrhythmia. *Sci Rep*. 2018;8:5251.

SUPPORTING INFORMATION

Additional Supporting Information may be found online in the Supporting Information section.

TABLE S1 Phantom T_1 , T_2 , and T_2^* times for SATURN, the conventional cardiac mapping sequences (SASHA, T_2 -prepared bSSFP, multi-GRE) and the reference SE and GRE methods for all single tubes

TABLE S2 T_1 , T_2 , and T_2^* times for the in vivo measurements for SATURN compared with the conventional cardiac mapping sequences (SASHA, T_2 -prepared bSSFP, multi-GRE) across all healthy subjects. Per-subject relaxation times are summarized as means and within-segment standard deviation, as highlighted in blue. The corresponding P-values for the t-test with Bonferroni correction are shown below

FIGURE S1 Simulations for the proposed SATURN sequence for varying T_1 (left), T_2 (middle), and T_2^* (right) for

different sources of error as (A) the rest period before the T_2 preparations, (B) Rician noise on the signal with corresponding SNR, (C) the heart rate in beats-per-minute (bpm) and (D) the T_2 preparation efficiency as a scale factor of the flip down and flip up 90° pulses of the T_2 preparation module. The relative deviation between simulated and true quantitative measures is depicted for each source of error. All simulations are performed with the common parameters (rest period of 10 seconds, noise-free, heart rate of 60 bpm, and T_2 preparation efficiency in %) and only the source of error was varied. In A, only deviations in T_2 are observed for a rest period of shorter than 5 seconds. In B, major deviations are observed for T_2^* dependent on the Rician noise. T_2 is less impacted and T_1 only slightly. C, no effect in neither T_1 , T_2 , and T_2^* was observed dependent on the heart rate. Deviations would be assumed for T_1 only if noise was added. In D, a strong drop in T_2 is observed for a decreased T_2 preparation efficiency

FIGURE S2 A, In vivo T_1 , T_2 , and T_2^* maps acquired with single-parameter reference methods (left) and the proposed SATURN sequence (right) for 2 healthy subjects. Visually homogeneous mapping is achieved throughout the myocardium for T_1 and T_2 , minor artifacts appear in T_2^* maps. Image quality appears visually comparable to the reference methods. B, Below the standard deviation (SD) maps are shown for the 3 relaxation times and the same subjects for SATURN and the reference methods

FIGURE S3 T_1 , T_2 , T_2^* , and R_2^* maps are shown for the acquisition with acceleration factor $R = 3$ (left), $R = 4$ (middle) and for $R = 4$ with additional regularization using SPIRiT + locally low rank (LLR) regularization (right). Quantitative measures with the standard deviation (shaded area) extracted from the SD maps along the myocardial wall are shown on the right side for $R = 3$ (blue), $R = 4$ (orange), and SPIRiT + LLR (yellow). Visual image quality is improved and precision is regained after the use of SPIRiT + LLR for $R = 4$. The color bar and the y-axis of the plot have the same ranges. The

corresponding quantitative times for the pixel-wise curve are windowed the same as the color bar left of the axis. The LLR algorithm takes around 200 seconds on a single core

FIGURE S4 In vivo T_2 times acquired with the T_2 -prepared bSSFP using a 3-parameter fit model and 4 dynamics and a 2-parameter fit model without the fourth dynamic (saturation). A, On the left side the mean T_2 times per healthy subject are correlated between the 2-parameter fit model and the 3-parameter fit model. B, On the right side, the Bland-Altman plot between 2 and 3 parameters is shown with a significant difference and an average bias of 5.85 ms increased T_2 when using the 2-parameter model. C, The representative T_2 maps are depicted with the corresponding bullseye plots (D) showing the within-segment mean and within-segment standard deviation across all subjects

FIGURE S5 A, Native T_1 , T_2 , and T_2^* maps without (top) and with motion correction using rigid registration (bottom). B, Magnitude images which indicate the difference encoded in blue and red between 2 images and the corresponding registered images below. On the left side contrast number 5 (max T_S^{\max}) was motion corrupted as also seen in the resulting T_1 map above. In the center image contrast number 3 (second T_2 preparation) was corrupted and on the right image along the different gradient echoes small translation was corrected. C, Signal intensity for a region of interest in the septal myocardium across 3 repetitions of the SATURN sequence

How to cite this article: Hermann J, Kellman P, Demirel OB, Akçakaya M, Schad LR, Weingärtner S. Free-breathing simultaneous T_1 , T_2 , and T_2^* quantification in the myocardium. *Magn Reson Med.* 2021;00:1–15. <https://doi.org/10.1002/mrm.28753>